

## Test Points

Seq.	Description	Nominal	Tolerance	Unit Measure	As Found	As Left	Uncertainty
1	CAL FACTOR FATc	910	9999 0		880	880	N/A
2	CW TORQUE 50RPM	0.50	0.53 0.47	N·m	0.49	0.49	0.011 N·m
3	CW TORQUE 100RPM	2.00	2.10 1.90	N·m	1.98	1.98	0.032 N·m
4	CW TORQUE 100RPM	5.00	5.15 4.85	N·m	4.97	4.97	0.074 N·m
5	CW TORQUE 100RPM	7.50	7.73 7.27	N·m	7.46	7.46	0.11 N·m
6	CW TORQUE 100RPM	10.00	10.30 9.70	N·m	10.16	10.16	0.14 N·m
7	CW TORQUE 100RPM	12.00	12.36 11.64	N·m	12.04	12.04	0.17 N·m
8	CW TORQUE 100RPM	15.00	15.45 14.55	N·m	14.94	14.94	0.21 N·m

End of Data

Note: A = Adjusted F = Failed L = Limited N/A = Outside Scope of Accreditation